

NBS
PUBLICATIONS

NATL INST OF STAND & TECH



A11107 257755



NBS SPECIAL PUBLICATION 260-92

U.S. DEPARTMENT OF COMMERCE / National Bureau of Standards

Standard Reference Materials:

Calibrated Glass Standards for Fission Track Use

QC
100
U57
No. 260-92
1934
c. 2

T

he National Bureau of Standards¹ was established by an act of Congress on March 3, 1901. The Bureau's overall goal is to strengthen and advance the nation's science and technology and facilitate their effective application for public benefit. To this end, the Bureau conducts research and provides: (1) a basis for the nation's physical measurement system, (2) scientific and technological services for industry and government, (3) a technical basis for equity in trade, and (4) technical services to promote public safety. The Bureau's technical work is performed by the National Measurement Laboratory, the National Engineering Laboratory, the Institute for Computer Sciences and Technology, and the Center for Materials Science.

The National Measurement Laboratory

Provides the national system of physical and chemical measurement; coordinates the system with measurement systems of other nations and furnishes essential services leading to accurate and uniform physical and chemical measurement throughout the Nation's scientific community, industry, and commerce; provides advisory and research services to other Government agencies; conducts physical and chemical research; develops, produces, and distributes Standard Reference Materials; and provides calibration services. The Laboratory consists of the following centers:

- Basic Standards²
- Radiation Research
- Chemical Physics
- Analytical Chemistry

The National Engineering Laboratory

Provides technology and technical services to the public and private sectors to address national needs and to solve national problems; conducts research in engineering and applied science in support of these efforts; builds and maintains competence in the necessary disciplines required to carry out this research and technical service; develops engineering data and measurement capabilities; provides engineering measurement traceability services; develops test methods and proposes engineering standards and code changes; develops and proposes new engineering practices; and develops and improves mechanisms to transfer results of its research to the ultimate user. The Laboratory consists of the following centers:

- Applied Mathematics
- Electronics and Electrical Engineering²
- Manufacturing Engineering
- Building Technology
- Fire Research
- Chemical Engineering²

The Institute for Computer Sciences and Technology

Conducts research and provides scientific and technical services to aid Federal agencies in the selection, acquisition, application, and use of computer technology to improve effectiveness and economy in Government operations in accordance with Public Law 89-306 (40 U.S.C. 759), relevant Executive Orders, and other directives; carries out this mission by managing the Federal Information Processing Standards Program, developing Federal ADP standards guidelines, and managing Federal participation in ADP voluntary standardization activities; provides scientific and technological advisory services and assistance to Federal agencies; and provides the technical foundation for computer-related policies of the Federal Government. The Institute consists of the following centers:

- Programming Science and Technology
- Computer Systems Engineering

The Center for Materials Science

Conducts research and provides measurements, data, standards, reference materials, quantitative understanding and other technical information fundamental to the processing, structure, properties and performance of materials; addresses the scientific basis for new advanced materials technologies; plans research around cross-country scientific themes such as nondestructive evaluation and phase diagram development; oversees Bureau-wide technical programs in nuclear reactor radiation research and nondestructive evaluation; and broadly disseminates generic technical information resulting from its programs. The Center consists of the following Divisions:

- Inorganic Materials
- Fracture and Deformation³
- Polymers
- Metallurgy
- Reactor Radiation

¹Headquarters and Laboratories at Gaithersburg, MD, unless otherwise noted; mailing address Gaithersburg, MD 20899.

²Some divisions within the center are located at Boulder, CO 80303.

³Located at Boulder, CO, with some elements at Gaithersburg, MD.

Standard Reference Materials:

NATIONAL BUREAU
OF STANDARDS
LIBRARY

QC100

. USPS

NO. 260-92

1984

C.2

Calibrated Glass Standards for Fission Track Use (Supplement to NBS SP 260-49)

B. Stephen Carpenter

National Measurement Laboratory
National Bureau of Standards
Gaithersburg, MD 20899



NBS Special publication

U.S. DEPARTMENT OF COMMERCE, Malcolm Baldrige, Secretary
NATIONAL BUREAU OF STANDARDS, Ernest Ambler, Director

Issued September 1984

Library of Congress Catalog Card Number: 84-601112

National Bureau of Standards Special Publication 260-92
Natl. Bur. Stand. (U.S.), Spec. Publ. 260-92, 21 pages (Sept. 1984)
CODEN: XNBSAV

U.S. GOVERNMENT PRINTING OFFICE
WASHINGTON: 1984

For sale by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402

PREFACE

Standard Reference Materials (SRM's) as defined by the National Bureau of Standards are "well-characterized materials, produced in quantity, that calibrate a measurement system to assure compatibility of measurement in the Nation." SRM's are widely used as primary standards in many diverse fields of science, industry, and technology, both within the United States and throughout the world. For many of the Nation's scientists and technologists it is of more than passing interest to know the measurements obtained and methods used by the analytical community when analyzing SRM's. An NBS series of papers, of which this publication is a member, called the NBS Special Publication - 260 Series is reserved for this purpose.

This 260 Series is dedicated to the dissemination of information on all phases of the preparation, measurement, and certification of NBS SRM's. In general, more detail will be found in these papers than is generally allowed, or desirable, in scientific journal articles. This enables the user to assess the validity and accuracy of the measurement processes employed, to judge the statistical analysis, and to learn details of techniques and methods utilized for work entailing the greatest care and accuracy. It is also hoped that these papers will provide sufficient additional information not found on the certificate so that new applications in diverse fields not foreseen at the time the SRM was originally issued will be sought and found.

Inquiries concerning the technical content of this paper should be directed to the author. Other questions concerned with the availability, delivery, price of specific SRM's should be addressed to:

Office of Standard Reference Materials
National Bureau of Standards
Gaithersburg, MD 20899

Stanley D. Rasberry, Chief
Office of Standard Reference Materials

OTHER NBS PUBLICATIONS IN THIS SERIES

- Catalog of NBS Standard Reference Materials (1984-85 edition), Catherine H. Hudson, ed., NBS Spec. Publ. 260 (February 1984).
- Michaelis, R. E., and Wyman, L. L. Standard Reference Materials: Preparation of White Cast Iron Spectrochemical Standards. NBS Misc. Publ. 260-1 (June 1964). COM74-11061**
- Michaelis, R. E., Wyman, L. L., and Flitsch, R., Standard Reference Materials: Preparation of NBS Copper-Base Spectrochemical Standards. NBS Misc. Publ. 260-2 (October 1964). COM74-11063**
- Michaelis, R. E., Yakowitz, H., and Moore, G. A., Standard Reference Materials: Metallographic Characterization of an NBS Spectrometric Low-Alloy Steel Standard. NBS Misc. Publ. 260-3 (October 1964). COM74-11060**
- Hague, J. L., Mears, T. W., and Michaelis, R. E., Standard Reference Materials: Sources of Information, NBS Misc. Publ. 260-4 (February 1965). COM74-11059
- Alvarez, R., and Flitsch, R., Standard Reference Materials: Accuracy of Solution X-Ray Spectrometric Analysis of Copper-Base Alloys. NBS Misc. Publ. 260-5 (March 1965). PB168068**
- Shultz, J. I., Standard Reference Materials: Methods for the Chemical Analysis of White Cast Iron Standards, NBS Misc. Publ. 260-6 (July 1965). COM74-11068**
- Bell, R. K., Standard Reference Materials: Methods for the Chemical Analysis of NBS Copper-Base Spectrochemical Standards. NBS Misc. Publ. 260-7 (October 1965). COM74-11067**
- Richmond, M.S., Standard Reference Materials: Analysis of Uranium Concentrates at the National Bureau of Standards. NBS Misc. Publ. 260-8 (December 1965). COM74-11066**
- Anspach, S. C., Cavallo, L. M., Garfinkel, S. B., Hutchinson, J. M. R., and Smith, C. N., Standard Reference Materials: Half Lives of Materials Used in the Preparation of Standard Reference Materials of Nineteen Radioactive Nuclides Issued by the National Bureau of Standards. NBS Misc. Publ. 260-9 (November 1965). COM74-11065**
- Yakowitz, H., Vieth, D. L., Heinrich, K. F. J., and Michaelis, R. E., Standard Reference Materials: Homogeneity Characterization of NBS Spectrometric Standards II: Cartridge Brass and Low-Alloy Steel, NBS Misc. Publ. 260-10 (December 1965). COM74-11064**
- Napolitano, A., and Hawkins, E. G., Standard Reference Materials: Viscosity of Standard Lead-Silica Glass, NBS Misc. Publ. 260-11 (November 1966). NBS Misc. Publ. 260-11**
- Yakowitz, H., Vieth, D. L., and Michaelis, R. E., Standard Reference Materials: Homogeneity Characterization of NBS Spectrometric Standards III: White Cast Iron and Stainless Steel Powder Compact, NBS Misc. Publ. 260-12 (September 1966). NBS Misc. Publ. 260-12**
- Spijkerman, J. L., Snediker, D. K., Ruegg, F. C., and DeVoe, J. R., Standard Reference Materials: Mossbauer Spectroscopy Standard for the Chemical Shift of Iron Compounds, NBS Misc. Publ. 260-13 (July 1967). NBS Misc. Publ. 260-13**
- Menis, O., and Sterling, J. T., Standard Reference Materials: Determination of Oxygen in Ferrrous Materials - SRM 1090, 1091, and 1092, NBS Misc. Publ. 260-14 (September 1966). NBS Misc. Publ. 260-14**
- Passaglia, E., and Shouse, P. J., Standard Reference Materials: Recommended Method of Use of Standard Light-Sensitive Paper for Calibrating Carbon Arcs Used in Testing Textiles for Colorfastness to Light, NBS Misc. Publ. 260-15 (June 1967). (Replaced by NBS Spec. Publ. 260-41.)
- Yakowitz, H., Michaelis, R. E., and Vieth, D. L., Standard Reference Materials: Homogeneity Characterization of NBS Spectrometric Standards IV: Preparation and Microprobe Characterization of W-20% MO Alloy Fabricated by Powder Metallurgical Methods, NBS Spec. Publ. 260-16 (January 1969). COM74-11062**
- Catanzaro, E. J., Champion, C. E., Garner, E. L., Marinenko, G., Sappenfield, K. M., and Shields, W. R., Standard Reference Materials: Boric Acid; Isotopic and Assay Standard Reference Materials, NBS Spec. Publ. 260-17 (February 1970). Out of Print.
- Geller, S. B., Mantek, P.A., and Cleveland, N. G., Standard Reference Materials: Calibration of NBS Secondary Standard Magnetic Tape (Computer Amplitude Reference) Using the Reference Tape Amplitude Measurement "Process A," NBS Spec. Publ. 260-18 (November 1969). (See NBS Spec. Publ. 260-29.)
- Paule, R. C., and Mandel, J., Standard Reference Materials: Analysis of Interlaboratory Measurements on the Vapor Pressure of Gold (Certification of Standard Reference Material 745). NBS Spec. Publ. 260-19 (January 1970). PB190071**
- Paule, R. C., and Mandel, J., Standard Reference Materials: Analysis of Interlaboratory Measurements on the Vapor Pressures of Cadmium and Silver, NBS Spec. Publ. 260-21 (January 1971). COM74-11359**
- Yakowitz, H., Fiori, C. E., and Michaelis, R. E., Standard Reference Materials: Homogeneity Characterization of Fe-3 Si Alloy, NBS Spec. Publ. 260-22 (February 1971). COM74-11357**
- Napolitano, A., and Hawkins, E. G., Standard Reference Materials: Viscosity of a Standard Borosilicate Glass, NBS Spec. Publ. 260-23 (December 1970). COM71-00157**
- Sappenfield, K. M., Marinenko, G., and Hague, J. L., Standard Reference Materials: Comparison of Redox Standards, NBS Spec. Publ. 260-24 (January 1972). COM72-50058**

- Hicho, G. E., Yakowitz, H., Rasberry, S. D., and Michaelis, R. E., Standard Reference Materials: A Standard Reference Material Containing Nominally Four Percent Austenite, NBS Spec. Publ. 260-25 (February 1971). COM74-11356**
- Martin, J. F., Standard Reference Materials: National Bureau of Standards-US Steel Corporation Joint Program for Determining Oxygen and Nitrogen in Steel, NBS Spec. Publ. 260-26 (February 1971). 85 cents* PB 81176620
- Garner, E. L., Machlan, L. A., and Shields, W. R., Standard Reference Materials: Uranium Isotopic Standard Reference Materials, NBS Spec. Publ. 260-27 (April 1971). COM74-11358**
- Heinrich, K. F. J., Myklebust, R. L., Rasberry, S. D., and Michaelis, R. E., Standard Reference Materials: Preparation and Evaluation of SRM's 481 and 482 Gold-Silver and Gold-Copper Alloys for Microanalysis, NBS Spec. Publ. 260-28 (August 1971). COM71-50365**
- Geller, S. B., Standard Reference Materials: Calibration of NBS Secondary Standard Magnetic Tape (Computer Amplitude Reference) Using the Reference Tape Amplitude Measurement "Process A-Model 2," NBS Spec. Publ. 260-29 (June 1971). COM71-50282
- Gorozhanina, R. S., Freedman, A. Y., and Shaevitch, A. B. (translated by M. C. Selby), Standard Reference Materials: Standard Samples Issued in the USSR (A Translation from the Russian), NBS Spec. Publ. 260-30 (June 1971). COM71-50283**
- Hust, J. G., and Sparks, L. L., Standard Reference Materials: Thermal Conductivity of Electrolytic Iron SRM 734 from 4 to 300 K, NBS Spec. Publ. 260-31 (November 1971). COM71-50563**
- Mavrodineanu, R., and Lazar, J. W., Standard Reference Materials: Standard Quartz Cuvettes, for High Accuracy Spectrophotometry, NBS Spec. Publ. 260-32 (December 1973). 55 cents* SN003-003-01213-1
- Wagner, H. L., Standard Reference Materials: Comparison of Original and Supplemental SRM 705, Narrow Molecular Weight Distribution Polystyrene, NBS Spec. Publ. 260-33 (May 1972). COM72-50526**
- Sparks, L. L., and Hust, J. G., Standard Reference Materials: Thermoelectric Voltage, NBS Spec. Publ. 260-34, (April 1972). COM72-50371**
- Sparks, L. L., and Hust, J. G., Standard Reference Materials: Thermal Conductivity of Austenitic Stainless Steel, SRM 735 from 5 to 280 K, NBS Spec. Publ. 260-35 (April 1972). 35 cents* COM72-50368**
- Cali, J. P., Mandel, J., Moore, L. J., and Young, D. S., Standard Reference Materials: A Reference Method for the Determination of Calcium in Serum, NBS SRM 915, NBS Spec. Publ. 260-36 (May 1972). COM72-50527**
- Shultz, J. I. Bell., R. K. Rains, T. C., and Menis, O., Standard Reference Materials: Methods of Analysis of NBS Clay Standards, NBS Spec. Publ. 260-37 (June 1972). COM72-50692**
- Richmond, J. C., and Hsia, J. J., Standard Reference Materials: Preparation and Calibration of Standards of Spectral Specular Reflectance, NBS Spec. Publ. 260-38 (May 1972). COM72-50528**
- Clark, A. F., Denson, V. A., Hust, J. G., and Powell, R. L., Standard Reference Materials: The Eddy Current Decay Method for Resistivity Characterization of High-Purity Metals, NBS Spec. Publ. 260-39 (May 1972). COM72-50529**
- McAdie, H. G., Garn, P. D., and Menis, O., Standard Reference Materials: Selection of Thermal Analysis Temperature Standards Through a Cooperative Study (SRM 758, 759, 760), NBS Spec. Publ. 260-40 (August 1972.). COM72-50776**
- Wood, L. A., and Shouse, P. J., Standard Reference Materials: Use of Standard Light-Sensitive Paper for Calibrating Carbon Arcs Used in Testing Textiles for Colorfastness to Light, NBS Spec. Publ. 260-41 (August 1972) COM72-50775**
- Wagner, H. L. and Verdier, P. H., eds., Standard Reference Materials: The Characterization of Linear Polyethylene, SRM 1475, NBS Spec. Publ. 260-42 (September 1972). COM72-50944**
- Yakowitz, H., Ruff, A. W., and Michaelis, R. E., Standard Reference Materials: Preparation and Homogeneity Characterization of an Austenitic Iron-Chromium-Nickel Alloy, NBS Spec. Publ. 260-43 (November 1972). COM73-50760**
- Schooley, J. F., Soulen, R. J., Jr., and Evans, G. A., Jr., Standard Reference Materials: Preparation and Use of Superconductive Fixed Point Devices, SRM 767, NBS Spec. Publ. 260-44 (December 1972). COM73-50037**
- Greifer, B., Maienthal, E. J., Rains, T. C., and Rasberry, S. D., Standard Reference Materials: Powdered Lead-Based Paint, SRM 1579, NBS Spec. Publ. 260-45 (March 1973). COM73-50226**
- Hust, J. G., and Giarratano, P. J., Standard Reference Materials: Thermal Conductivity and Electrical Resistivity Standard Reference Materials: Austenitic Stainless Steel, SRM's 735 and 798, from 4 to 1200 K, NBS Spec. Publ. 260-46 (March 1975). SN003-003-01278-5*
- Hust, J. G., Standard Reference Materials: Electrical Resistivity of Electrolytic Iron, SRM 797, and Austenitic Stainless Steel, SRM 798, from 5 to 280 K, NBS Spec. Publ. 260-47 (February 1974). COM74-50176**
- Mangum, B. W., and Wise, J. A., Standard Reference Materials: Description and Use of Precision Thermometers for the Clinical Laboratory, SRM 933 and SRM 934, NBS Spec. Publ. 260-48 (May 1974). 60 cents* SN003-003-01278-5
- Carpenter, B. S., and Reimer, G. M., Standard Reference Materials: Calibrated Glass Standards for Fission Track Use, NBS Spec. Publ. 260-49 (November 1974). COM74-51185

- Hust, J. G., and Giarratano, P. J., Standard Reference Materials: Thermal Conductivity and Electrical Resistivity Standard Reference Materials: Electrolytic Iron, SRM's 734 and 797 from 4 to 1000 K, NBS Spec. Publ. 260-50 (June 1975). \$1.00* SN003-003-01425-7
- Mavrodineanu, R., and Baldwin, J. R., Standard Reference Materials: Glass Filters As a Standard Reference Material for Spectrophotometry; Selection; Preparation; Certification; Use-SRM 930, NBS Spec. Publ. 260-51 (November 1975). \$1.90* SN003-003-01481-8
- Hust, J. G., and Giarratano, P. J., Standard Reference Materials: Thermal Conductivity and Electrical Resistivity Standard Reference Materials 730 and 799, from 4 to 3000 K, NBS Spec. Publ. 260-52 (September 1975). \$1.05* SN003-003-01464-8
- Durst, R. A., Standard Reference Materials: Standardization of pH Measurements, NBS Spec. Publ. 260-53 (December 1975, Revised). \$1.05 SN003-003-01551-2
- Burke, R. W., and Mavrodineanu, R., Standard Reference Materials: Certification and Use of Acidic Potassium Dichromate Solutions as an Ultraviolet Absorbance Standard, NBS Spec. Publ. 260-54 (August 1977). \$3.00* SN003-003-01828-7
- Ditmars, D. A., Cezairliyan, A., Ishihara, S., and Douglas, T. B., Standard Reference Materials: Enthalpy and Heat Capacity; Molybdenum SRM 781, from 273 to 2800 K, NBS Spec. Publ. 260-55 (September 1977). \$2.20* SN003-003-01836-8
- Powell, R. L., Sparks, L. L., and Hust, J. G., Standard Reference Materials: Standard Thermocouple Materials, Pt.67: SRM 1967, NBS Spec. Publ. 260-56 (February 1978). \$2.20* SN003-003-018864
- Call, J. P. and Plebanski, T., Guide to United States Reference Materials, NBS Spec. Publ. 260-57 (February 1978). \$2.20* PB 277173
- Barnes, J. D., and Martin, G. M., Standard Reference Materials: Polyester Film for Oxygen Gas Transmission Measurements SRM 1470, NBS Spec. Publ. 260-58 (June 1979) \$2.00* SN003-003-02077
- Chang, T., and Kahn, A. H. Standard Reference Materials: Electron Paramagnetic Resonance Intensity Standard; SRM 2601, NBS Spec. Publ. 260-59 (August 1978) \$2.30* SN003-003-01975-5
- Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., and Moody, J. R., Standard Reference Materials: A Reference Method for the Determination of Sodium in Serum, NBS Spec. Publ. 260-60 (August 1978). \$3.00* SN003-003-01978-0
- Verdier, P. H., and Wagner, H. L., Standard Reference Materials: The Characterization of Linear Polyethylene (SRM 1482, 1483, 1484), NBS Spec. Publ. 260-61 (December 1978). \$1.70* SN003-003-02006-1
- Soulen, R. J., and Dove, R. B., Standard Reference Materials: Temperature Reference Standard for Use Below 0.5 K (SRM 768). NBS Spec. Publ. 260-62 (April 1979). \$2.30* SN003-003-02047-8
- Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., Machlan, L. A., and Gramlich, J. W., Standard Reference Materials: A Reference Method for the Determination of Potassium in Serum. NBS Spec. Publ. 260-63 (May 1979). \$3.75* SN003-003-02068
- Velapoldi, R. A., and Mielenz, K. D., Standard Reference Materials: A Fluorescence Standard Reference Material Quinine Sulfate Dihydrate (SRM 936), NBS Spec. Publ. 260-64 (January 1980). \$4.25* SN003-003-02148-2
- Marinenko, R. B., Heinrich, K. F. J., and Ruegg, F. C., Standard Reference Materials: Micro-Homogeneity Studies of NBS Standard Reference Materials, NBS Research Materials, and Other Related Samples. NBS Spec. Publ. 260-65 (September 1979). \$3.50* SN003-003-02114-1
- Venable, W. H., Jr., and Eckerle, K. L., Standard Reference Materials: Didymium Glass Filters for Calibrating the Wavelength Scale of Spectrophotometers (SRM 2009, 2010, 2013). NBS Spec. Publ. 260-66 (October 1979). \$3.50* SN003-003-02127-0
- Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., Murphy, T. J., and Gramlich, J. W., Standard Reference Materials: A Reference Method for the Determination of Chloride in Serum, NBS Spec. Publ. 260-67 (November 1979). \$3.75* SN003-003-02136-9
- Mavrodineanu, R. and Baldwin, J. R., Standard Reference Materials: Metal-On-Quartz Filters as a Standard Reference Material for Spectrophotometry-SRM 2031, NBS Spec. Publ. 260-68 (April 1980). \$4.25* SN003-003-02167-9
- Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., Machlan, L. A., Garner, E. L., and Rains, T. C., Standard Reference Materials: A Reference Method for the Determination of Lithium in Serum, NBS Spec. Publ. 260-69 (July 1980). \$4.25* SN003-003-02214-4
- Marinenko, R. B., Biancaniello, F., Boyer, P. A., Ruff, A. W., DeRobertis, L., Standard Reference Materials: Preparation and Characterization of an Iron-Chromium-Nickel Alloy for Micro-analysis, NBS Spec. Publ. 260-70 (May 1981). \$2.50* SN003-003-02328-1
- Seward, R. W., and Mavrodineanu, R., Standard Reference Materials: Summary of the Clinical Laboratory Standards Issued by the National Bureau of Standards, NBS Spec. Publ. 260-71 (November 1981). \$6.50* SN003-003-02381-7
- Reeder, D.J., Coxon, B., Enagonio, D., Christensen, R. G., Schaffer, R., Howell, B. F., Paule, R. C., Mandel, J., Standard Reference Materials: SRM 900, Antiepilepsy Drug Level Assay Standard, NBS Spec. Publ. 260-72 (June 1981). \$4.25* SN003-003-02329-9

- Interrante, C. G., and Hicho, G. E., Standard Reference Materials: A Standard Reference Material Containing Nominally Fifteen Percent Austenite (SRM 486), NBS Spec. Publ. 260-73 (January 1982). \$2.75* SN003-003-02386-8
- Marinenko, R. B., Standard Reference Materials: Preparation and Characterization of K-411 and K-414 Mineral Glasses for Microanalysis: SRM 470. NBS Spec. Publ. 260-74 (April 1982). \$3.50 SN003-003-023-95-7
- Weidner, V. R., and Hsia, J. J., Standard Reference Materials: Preparation and Calibration of First Surface Aluminum Mirror Specular Reflectance Standards (SRM 2003a), NBS Spec. Publ. 260-75 (May 1982). \$3.75 SN003-003-023-99-0
- Hicho, G. E. and Eaton, E. E., Standard Reference Materials: A Standard Reference Material Containing Nominally Five Percent Austenite (SRM 485a), NBS Spec. Publ. 260-76 (August 1982). \$3.50 SN003-003-024-33-3
- Furukawa, G. T., Riddle, J. L., Bigge, W. G., and Pfeiffer, E. R., Standard Reference Materials: Application of Some Metal SRM's as Thermometric Fixed Points, NBS Spec. Publ. 260-77 (August 1982). \$6.00 SN003-003-024-34-1
- Hicho, G. E. and Eaton, E. E., Standard Reference Materials: Standard Reference Material Containing Nominally Thirty Percent Austenite (SRM 487), NBS Spec. Publ. 260-78 (September 1982). \$3.75* SN003-003-024-35-0
- Richmond, J. C., Hsia, J. J., Weidner, V. R., and Wilmering, D. B., Standard Reference Materials: Second Surface Mirror Standards of Specular Spectral Reflectance (SRM's 2023, 2024, 2025), NBS Spec. Publ. 260-79 (October 1982). \$4.50* SN003-003-024-47-3
- Schaffer, R., Mandel, J., Sun, T., Cohen, A., and Hertz, H. S., Standard Reference Materials: Evaluation by an ID/MS Method of the AACC Reference Method for Serum Glucose, NBS Spec. Publ. 260-80 (October 1982). \$4.75* SN003-003-024-43-1
- Burke, R. W., and Mavrodineanu, R. (NBS retired), Standard Reference Materials: Accuracy in Analytical Spectrophotometry, NBS Spec. Publ. 260-81 (April 1983). \$6.00* SN003-003-024-84-8
- Weidner, V. R., Standard Reference Materials: White Opal Glass Diffuse Spectral Reflectance Standards for the Visible Spectrum (SRM's 2015 and 2016), NBS Spec. Publ. 260-82 (April 1983). \$3.75* SN003-003-024-89-9**
- Bowers, G. N., Jr., Alvarez, R., Cali, J. P. (NBS retired), Eberhardt, K. R., Reeder, D. J., Schaffer, R., Uriano, G. A., Standard Reference Materials: The Measurement of the Catalytic (Activity) Concentration of Seven Enzymes in NBS Human Serum SRM 909, NBS Spec. Publ. 260-83 (June 1983). \$4.50* SN003-003-024-99-6
- Gills, T. E., Seward, R. W., Collins, R. J., and Webster, W. C., Standard Reference Materials: Sampling, Materials Handling, Processing, and Packaging of NBS Sulfur in Coal Standard Reference Materials, 2682, 2683, 2684, and 2685, NBS Spec. Publ. 260-84 (August 1983). \$4.50* SN003-003-025-20-8
- Swyt, D. A., Standard Reference Materials: A Look at Techniques for the Dimensional Calibration of Standard Microscopic Particles, NBS Spec. Publ. 260-85 (September 1983). \$5.50* SN003-003-025-21-6
- Hicho, G. E. and Eaton, E. E., Standard Reference Materials: A Standard Reference Material Containing Two and One-Half Percent Austenite, SRM 488, NBS Spec. Publ. 260-86 (December 1983). \$1.75* SN003-003-025-41-1
- Mangum, B. W., Standard Reference Materials: SRM 1969: Rubidium Triple-Point - A Temperature Reference Standard Near 39.30 °C, NBS Spec. Publ. 260-87 (December 1983). \$2.25* SN003-003-025-44-5
- Gladney, E. S., Burns, C. E., Perrin, D. R., Roelands, I., and Gills, T. E., Standard Reference 1982 Compilation of Elemental Concentration Data for NBS Biological, Geological, and Environmental Standard Reference Materials. NBS Spec. Publ. 260-88 (March 1984). SN003-003-02565-8
- Hust, J. G., Standard Reference Materials: A Fine-Grained, Isotropic Graphite for Use as NBS Thermophysical Property RM's from 5 to 2500 K, NBS Spec. Publ. 260-89 (In Press).
- Hust, J. G., and Lankford, A. B., Standard Reference Materials: Update of Thermal Conductivity and Electrical Resistivity of Electrolytic Iron, Tungsten, and Stainless Steel, NBS Spec. Publ. 260-90 (In Press).
- Goodrich, L. F., Vecchia, D. F., Pittman, E. S., Ekin, J. W., and Clark, A. F., Standard Reference Materials: Critical Current Measurements on an NbTi Superconducting Wire Standard Reference Material, NBS Spec. Publ. 260-91 (In Press).
- Carpenter, B. S., Standard Reference Materials: Calibrated Glass Standards for Fission Track Use (Supplement to NBS Spec. Publ. 260-49). NBS Spec. Publ. 260-92 (In Press).
- * Send order with remittance to Superintendent of Documents, US Government Printing Office Washington, DC 20402. Remittance from foreign countries should include an additional one-fourth of the purchase price for postage.
- ** May be ordered from: National Technical Information Services (NTIS). Springfield Virginia 22151.

CONTENTS

	PAGE
I. INTRODUCTION	1
II. CHARACTERIZATION OF THE GLASS STANDARDS	1
III. CERTIFICATION AND PACKAGE DESCRIPTION	1
IV. PREPARATION AND IRRADIATION PROCEDURES	2
V. FLUENCE DETERMINATION	2
VI. RECOMMENDED USE OF STANDARD REFERENCE MATERIALS	3
VII. ACKNOWLEDGEMENT	3
VIII. REFERENCES	4

LIST OF TABLES

<u>Table No.</u>		PAGE
I.	The Uranium Concentration ²³⁵ U Isotopic Abundance	5
II.	Concentration of Elements That May Cause Track Interferences	5
III.	Activation Energies for Track Fading in the NBS Standard Glasses	6
IV.	Approximate Track Densities	6
V.	SRM 962a Fission Track Glass Standard	7
VI.	SRM 963a Fission Track Glass Standard	8

LIST OF FIGURES

	PAGE
Figure I. Annealing studies of uranium fission tracks in the 50 ppm SRM 962a. Extrapolation of the data on this Arrhenius plot indicates that glasses used as track references should be stored in a cool place in order to prevent partial annealing of tracks.	9
Figure II. Position of RT-3 and RT-4 irradiated glasses and the four unirradiated glasses (UG) within SRM package.	10
Figure III. Arrangement of the glass, metal foils and detectors during neutron irradiation . .	11
Figure IV. Neutron cross sections for copper, gold, and cobalt as a function of neutron energy	12

CALIBRATED GLASS STANDARDS FOR FISSION TRACK USE-SUPPLEMENT

B. Stephen Carpenter

National Measurement Laboratory
National Bureau of Standards
Gaithersburg, MD 20899

Two glasses of different uranium concentrations were prepared and reissued for certification by the National Bureau of Standards as standards for use as neutron monitors to aid fission track studies. These Standard Reference Materials (SRM's) and their uranium concentrations are: SRM 962a (37.4 ppm) and SRM 963a (0.823 ppm). These glass wafers were irradiated in the National Bureau of Standards Research Reactor and the neutron fluence was monitored using copper and gold foils, as well as an iron-cobalt foil.

Key Words: Fission tracks; Flux monitors; glass standards; Standard Reference Material; thermal neutron irradiation; uranium.

I. INTRODUCTION

The fission track technique has become an accepted and widely used method for uranium analysis and radiometric dating in many laboratories. Therefore, the need exists for a primary standard to enable correlation of inter-laboratory results. In order to fulfill this need, the National Bureau of Standards (NBS), through the Office of Standard Reference Materials, has prepared and reissued for certification Fission Track Glass Standards, SRM's 962a and 963a,† (1,2). Although these are not the first standards prepared for fission track work (2,3,4), they are the first readily available fission track standards for monitoring neutron fluence during irradiations and for comparing track densities. Since the glasses have some inadequacies which will be discussed, they are not the "ideal" standards; they are, however, the best standards available at this time.

II. CHARACTERIZATION OF THE GLASS STANDARDS

The Fission Track Glass Standards, SRM numbers 962a and 963a are available at two different uranium concentration levels, 37.38 ppm and 0.823 ppm, respectively. These glass wafers, 12 mm in diameter and 2 mm thick, with a nominal base composition of 72% SiO₂, 12% CaO, 14% Na₂O, and 2% Al₂O₃, came from the same lot of material as the Trace Element in Glass SRM's 612 and 614 (5,6,7). This series of SRM's has been analyzed for homogeneity and was certified for approximately half of the more than 60 trace elements present at the four concentration levels, including uranium. However, in the preparation of these glasses, the uranium used for the dopant was depleted uranium, and each concentration has a different uranium-235 isotopic abundance (as seen in Table I). In determining uranium by the fission track technique, thorium and boron are two major elements that could cause possible interference since they emit charged particles that may be recorded in many external detectors. The concentrations of these two elements are given in Table II. In the certification of these glasses, the analytical competences employed by the Center for Analytical Chemistry were isotope dilution mass spectrometry, polarography, flame emission and atomic absorption spectrometry, spectrophotometry, and nuclear methods of analysis, which included both neutron activation analysis and nuclear track technique.

To further characterize these standards, the thermal stability of fission tracks in the glasses was also studied to determine the fading rate of the tracks when subjected to heat. This thermal stability was determined by the activation energy which is the amount of energy needed for track annealing. Reimer et al (8) found that the glasses had relatively low thermal stability. Extrapolation of track fading data indicates that tracks could begin to decrease at 50 °C after a time of only several months. Figure I illustrates this extrapolation for SRM 962. Table III lists the activation energies calculated from the track fading experiments that would produce a specific degree of track reduction. It is recommended that these glasses be stored below 20 °C when not being used.

III. CERTIFICATION AND PACKAGE DESCRIPTION

The glasses are certified for the total neutron fluence ($n \cdot \text{cm}^{-2}$), in which they were irradiated. Copper and gold foils, as well as an iron-cobalt foil, were chosen to monitor the neutron fluence since many laboratories routinely use these foils.

The SRM package contains four unirradiated glass wafers and two separately irradiated wafers. The irradiations were performed in either of two positions of the NBS Research Reactor, each position providing different neutron energy spectra (9,10). Each package has two numbers designating the specific irradiated glasses enclosed. Figure II illustrates the configuration of the SRM package.

† Available from Office of Standard Reference Materials, National Bureau of Standards, Gaithersburg, MD 20899.

IV. PREPARATION AND IRRADIATION PROCEDURES

All of the glasses were ground flat and polished on both sides, cleaned in dilute nitric acid solution (1 part HNO₃: 10 parts H₂O) for 2 minutes, rinsed in distilled water and in ethanol, and air dried. Each cleaned wafer was then sandwiched between two pre-cleaned numbered detectors, Lexan* polycarbonate and muscovite mica. The combination was placed in a polyethylene bag which was hermetically vacuum sealed to ensure intimate contact between detectors and wafers. Each package was then placed in individual irradiation rabbits.

Pre-weighed foil flux monitors of copper, 0.1270-mm thick with average weight 0.021 g; gold, 0.0254-mm thick with average weight 0.031 g; iron-cobalt, 0.0254-mm thick with average weight 0.0037 g, were placed in every tenth rabbit starting with the first rabbit. (See Figure III for sample configuration during irradiation.) Each glass concentration level was divided into two groups of 128 samples, and each sample was irradiated individually in one of two different pneumatic tubes of the NBS Reactor (NBSR). In the irradiation positions used for this work, RT-3 has a cadmium ratio of 10.2 for gold foils and 65 for copper foils, while RT-4 has a cadmium ratio of 87 for gold and 536 for copper foils (10). The irradiation time used depended on the uranium concentration and uranium-235 isotopic abundance (see Table II). Due to the NBSR operating schedule the 962a series of glasses were not irradiated at the usual 10 MW Reactor power but at a reduced power or 8.5 MW.

V. FLUENCE DETERMINATION

After irradiation, the gross radioactivity was allowed to decay and the foil flux monitors were removed from the irradiation containers while external detectors and glasses were separated. The activity of the 0.511 MeV, ⁶⁴Cu positron and its 1.345 MeV gamma-ray was counted with a lithium drifted germanium detector; the same procedure was followed for counting the activity of the 0.412 MeV gamma-ray of ¹⁹⁸Au, the 1.099 and 1.292 MeV gamma-rays of Fe and the 1.173 and 1.332 MeV gamma-rays of ⁶⁰Co. Each foil was counted at various distances from the detector. The detectors had been previously calibrated with NBS SRM's chosen from the NBS radioactivity series; therefore, absolute peak efficiencies were determined for the various γ -ray energies of copper, gold, iron and cobalt. The copper foils were placed between aluminum discs 0.32-mm thick in order to ensure annihilation of all ⁶⁴Cu, 0.511-MeV positrons close to the foil. After counting all the foils that had been irradiated in the two reactor positions, the neutron flux for each foil was determined using nuclear and physical constants. The following equations were used to calculate the thermal neutron flux:

$$\text{Activity } A = \frac{C}{\Xi} a \quad (1)$$

A = Activity in nuclear transformations per second at T₀ (Time of zero decay, i.e., and of irradiation)

C = Counts per second at T₀

Ξ = Absolute counter efficiency for the appropriate energy gamma-ray

a = Gamma ray intensity; ⁶⁴Cu(0.511 MeV) = 0.382 and (1.345 MeV) = 0.0062;

¹⁹⁸Au(0.412 MeV) = 0.955 ⁵⁹Fe(1.099 MeV) = 0.565 and (1.292 MeV)

= 0.435; and ⁶⁰Co(1.173 MeV) = 0.999 and (1.332 MeV) = 0.998 (11)

$$\phi = \frac{A}{\sigma_{th} N (1 - e^{-\lambda t})} \quad (2)$$

ϕ = Thermal neutron flux (n.cm⁻².sec⁻¹)

σ_{th} = Thermal neutron cross section; ⁶³Cu = 4.5x10⁻²⁴ cm²,
¹⁹⁷Au = 9.8.8x10⁻²⁴ cm², ⁵⁸Fe = 1.4x10⁻²⁴ cm² and ⁵⁹Co = 18.2x10⁻²⁴ cm² (Figure IV)

N = Total number of target nuclei

λ = Disintegration constant of radionuclide (⁶⁴Cu, ¹⁹⁸Au, ⁵⁹Fe or ⁶⁰Co material)

t = Time duration of exposure to neutron field

then the neutron fluence is

$$\Phi = \phi t \quad (3)$$

The external detectors, previously separated from the glass wafers, were chemically etched to optically reveal the fission tracks. The polycarbonate detectors were etched in a 6.5N NaOH solution for 45 minutes at a constant temperature of 50 °C, then rinsed in distilled water and air dried.

The mica detectors were etched in 48% HF for 45 minutes at 25 °C, rinsed in water, and heated to dry and to volatilize any remaining HF. The irradiated glass wafers were etched in 24% HF for 20 seconds at 25 °C, rinsed with water for 15 minutes and allowed to air dry.

After etching, both of the external detectors were counted with the aid of an image analyzing microscope under the pre-established conditions of either a minimum of at least 1,000 tracks or 50 random fields of view. The random fields of view were of an area of $1.74 \times 10^{-4} \text{ cm}^2$ at a magnification of 50X. An example of the tracks densities obtained from the counting of the external detectors and glass wafers are given in Table IV. The same counting criteria were employed for the glasses. Counting data was also collected for each series of metal foils, the two external detectors and both sides of the glasses, and a statistical evaluation was performed for trends or variations in the neutron flux.

The fluence determinations for glasses SRM 962a and 963a are reported in Tables V and VI, respectively. The neutron fluence was determined by multiplying the flux by the irradiation time. There were no obvious trends or variations outside of those imposed by irradiated material.

VI. RECOMMENDED USE OF STANDARD REFERENCE MATERIALS

The SRM's were prepared and certified in a manner to minimize restriction of their specific use. The selection of a reactor facility, the etching conditions, and the counting criteria are all the choice of the individual laboratory.

In order to determine an unknown neutron fluence with this set of glasses, the following procedures are suggested.

1. Determine whether the laboratory will be using glass, mica, or polycarbonate as its primary counting detector medium.
2. Place a piece, or pieces, of unirradiated SRM 962a or SRM 963a with, or without, an external detector (mica or polycarbonate) in the irradiation container with samples.
3. After irradiation, polish all pieces of glass, including RT-3 and RT-4 glasses (SRM 962a or SRM 963a irradiated glasses) to reveal an internal surface. At least 30 μm should be removed. Simultaneously etch all pieces of glass, rinse well, and count. The unknown neutron fluence can then be calculated from the resulting track densities. Repeat the polishing and etching of RT-3 and RT-4 glasses each time a set of samples are irradiated and counted.
4. Etch the external detectors which were over the RT-3 and RT-4 SRM glasses during irradiation. (Note these detectors can only be etched once). At the same time, etch the external detector used in step 2 when unirradiated pieces of SRM 962a or SRM 963a were prepared for irradiation. Then count both sets of detectors and determine the neutron fluence from the track densities.

The users of this material are cautioned that the track densities shown in Table IV are approximate. This is because the observed track density, particularly in glass, is dependent upon etching conditions, magnification, and microscope illumination. It is also suggested that all of the irradiated SRM 962a and SRM 963a glasses be stored below 20 °C to reduce the chance of track annealing.

The two SRM irradiated glasses allow an additional option. There was a difference in the thermal to fast neutron ratio for the irradiation of these wafers, as indicated by the cadmium ratios for gold and copper foil. The wafer selected for the counting comparisons should most closely approximate the neutron energy conditions of the reactor used. The specific glass used as a standard should always be mentioned by its NBS-SRM number in publication.

ACKNOWLEDGEMENT

The author would like to thank the members of the Reactor Operation Section and T. Mitlehner for their help in the irradiation of the glass wafers, and also to T. Mitlehner for her help in packaging and preparation of these SRMs. Both appreciation and thanks are extended to C. W. Naeser, N. Naeser, M. Miyachi, R. Zimmermann and K. Crowley all of the U.S. Geological Survey, Denver, Colo., for their assistance in counting the two external detectors as well as the irradiated glasses. Appreciation is also extended to R. Paule, Statistical Consultant for the National Measurement Laboratory, for his statistical evaluation. A special note of thanks is extended to T. E. Gills of the Office of Standard Reference Materials

Disclaimer

Certain commercial equipment, instruments, or materials are identified in this paper in order to adequately specify the experimental procedure. In no case does such identification imply recommendation or endorsement by the National Bureau of Standards, nor does it imply that the material or equipment identified is necessarily the best available for the purpose.

REFERENCES

1. National Bureau of Standards Certificates of Analysis for SRM's 961 through 964.
2. NBS Special Publication 260-49 Standard Reference Materials: Calibrated Glass Standards for Fission Track Use, SRM 961, SRM 962, SRM 963 and SRM 964, November 1974.
3. Schreurs, J. W., Friedman, A. M., Rokop, D. J., Hair, M. W., and Walker, R. M., *Radiation Effects*, 7, 231 (1971).
4. Fleischer, R. L., Price, P. B., and Walker, R. M., *Nuclear Science and Engineering*, 22, 153 (1965).
5. National Bureau of Standards Certificates of Analysis for SRM's 610 through 616.
6. Carpenter, B. S., *Anal. Chem.* 44, 600 (1972).
7. Barnes, I. L., Garner, E. L., Gramlich, J. W., Moore, L. J. Murphy, T. J., Machlan, L. A., and Shields, W. R., *Anal. Chem.*, 45, 880 (1973).
8. Reimer, G. M., Wagner, G. A., and Carpenter, B. S., *Radiation Effects*, 15, 273 (1972).
9. LaFleur, P. D. and Becker, D. A., editors, National Bureau of Standards (U. S.), Technical Note 548, Washington, D.C. (1970).
10. Becker, D. A. and LaFleur, P. D., *Journal of Radioanalytical Chemistry*, 19, 149 (1974).
11. Lederer, C. M., Hollander, J. M., and Perlman, I., Table of Isotopes, Seventh edition, John Wiley and Sons, Inc. (1978).

Table I
The Uranium Concentration ²³⁵U
Isotopic Abundance

Total U Concentration ppm (by weight)	²³⁵ U Atom Percent	Reactor Location and Irradiation Time (s)
SRM 962a 37.38 ± 0.08	0.2392	RT-3 9.2 RT-4 34.5
SRM 963a 0.823 ± 0.002	0.2792	RT-3 80 RT-4 300

Table II
Concentration of Elements
That May Cause Track Interferences

	Boron (ppm by weight)	Thorium (ppm by weight)
SRM 962a	32	37.79 ± 0.08
SRM 963a	1.3	0.748 ± 0.006

Table III
Activation Energies for Track Fading
in the NBS Standard Glasses (8)

Density Reduction ρ/ρ_0	SRM 964	SRM 963	SRM 962	SRM 961
0.95	0.8 eV	0.7 eV	1.2 eV	0.8 eV
0.90	1.0	0.8	1.3	1.0
0.80	1.3	0.9	1.5	1.2
0.70	1.5	1.1	1.7	1.4
0.60	1.7	1.2	1.8	1.6
0.50	1.8	1.5	2.0	1.7

Table IV
Approximate Track Densities

SRM 962a

	Glass ¹	Mica ²	Polycarbonate ³
RT-3	10 x 10 ⁴ t·cm ⁻²	9 x 10 ⁴ t·cm ⁻²	9 x 10 ⁴ t·cm ⁻²
RT-4	9 x 10 ⁴ t·cm ⁻²	8 x 10 ⁴ t·cm ⁻²	8 x 10 ⁴ t·cm ⁻²

SRM 963a

	Glass ¹	Mica ²	Polycarbonate ³
RT-3	2.5 x 10 ⁴ t·cm ⁻²	2.2 x 10 ⁴ t·cm ⁻²	2.2 x 10 ⁴ t·cm ⁻²
RT-4	2.4 x 10 ⁴ t·cm ⁻²	2.2 x 10 ⁴ t·cm ⁻²	2.1 x 10 ⁴ t·cm ⁻²

- 1) Polished internal surface etched for 20 seconds in 24% HF at 25° C, counted at 500X.
- 2) External (2 π) surface etched for 45 minutes in 48% HF at 25° C, counted at 500X.
- 3) External (2 π) surface etched for 45 minutes in 6.5 N NaOH at 50° C, counted at 500X.

Table V
SRM 962a Fission Track Glass Standard
Certified Values

NBS Reactor Position	Neutron Fluence Mean Value and Standard Deviation ($\times 10^{14} \text{ n} \cdot \text{cm}^{-2}$) ^{a, b}	Tolerance Interval ^c
<u>Cu Foil</u>		
RT-4	3.87 \pm 0.07	\pm 0.29
RT-3	4.37 \pm 0.09	\pm 0.35
<u>Au Foil</u>		
RT-4	4.17 \pm 0.08	\pm 0.33
RT-3	4.75 \pm 0.05	\pm 0.19
"For Information Purposes Only" ^d		
<u>Fe Foil</u>		
RT-4	3.89 \pm 0.08	\pm 0.31
RT-3	4.25 \pm 0.16	\pm 0.64
<u>Co Foil</u>		
RT-4	3.79 \pm 0.12	\pm 0.48
RT-3	4.49 \pm 0.19	\pm 0.75

^aStandard Deviations refer to individual metal foils.

^bIrradiations were performed at a power of 8.5 megawatts; 9.2 seconds in RT-3, or 34.5 seconds in RT-4.

^cTolerance intervals for 95% of the SRM population at the 95% confidence level.

^dThese results show greater imprecision.

Table VI
SRM 963a Fission Track Glass Standard
Certified Values

NBS Reactor Position	Neutron Fluence Mean Values and Standard Deviation ($\times 10^{14} \text{ n} \cdot \text{cm}^{-2}$) ^{a, b}	Tolerance Interval ^c
<u>Cu Foil</u>		
RT-4	39.5 \pm 0.1	\pm 0.5
RT-3	41.2 \pm 0.7	\pm 2.6
<u>Au Foil</u>		
RT-4	43.0 \pm 0.7	\pm 2.9
RT-3	45.8 \pm 0.4	\pm 1.5
"For Information Purposes Only" ^d		
<u>Fe Foil</u>		
RT-4	40.2 \pm 1.3	\pm 5.3
RT-3	40.7 \pm 1.1	\pm 4.5
<u>Co Foil</u>		
RT-4	41.2 \pm 1.7	\pm 6.7
RT-3	45.7 \pm 2.2	\pm 8.7


^aStandard Deviations refer to individual metal foils.

^bIrradiations were performed at a power of 10.0 megawatts; 80 seconds in RT-3, or 300 seconds in RT-4.

^cTolerance intervals for 95% of the SRM population at the 95% confidence level.

^dThese results show greater imprecision.

Figure I. Annealing Studies Indicating the Fraction of Uranium Fission Tracks Reduced as a Function of Time and Temperature in the 50 ppm SRM 962



**Figure II. Position of RT-3 and RT-4
Irradiated Glasses**




Figure III. Arrangement of the Glass, Metal Foils and Detectors During Neutron Irradiations





Figure IV. Neutron Cross Sections of Copper, Gold and Cobalt as a Function of Neutron Energy



U.S. DEPT. OF COMM. BIBLIOGRAPHIC DATA SHEET (See instructions)	1. PUBLICATION OR REPORT NO. NBS/SP-260/92	2. Performing Organ. Report No.	3. Publication Date Sept. 1984
4. TITLE AND SUBTITLE Standard Reference Materials: Calibrated Glass Standards for Fission Track Use (Supplement to NBS Spec. Publ. 260-49)			
5. AUTHOR(S) B. S. Carpenter			
6. PERFORMING ORGANIZATION (If joint or other than NBS, see instructions) NATIONAL BUREAU OF STANDARDS DEPARTMENT OF COMMERCE WASHINGTON, D.C. 20234		7. Contract/Grant No.	8. Type of Report & Period Covered Final
9. SPONSORING ORGANIZATION NAME AND COMPLETE ADDRESS (Street, City, State, ZIP) Same as No. 6			
10. SUPPLEMENTARY NOTES Library of Congress Catalog Card Number: 84-601112 <input type="checkbox"/> Document describes a computer program; SF-185, FIPS Software Summary, is attached.			
11. ABSTRACT (A 200-word or less factual summary of most significant information. If document includes a significant bibliography or literature survey, mention it here) <p>Two glasses of different uranium concentrations were prepared and reissued for certification by the National Bureau of Standards as standards for use as neutron monitors to aid fission track studies. These Standard Reference Materials (SRM's) and their uranium concentrations are: SRM 962a (37.4 ppm) and SRM 963a (0.823 ppm). These glass wafers were irradiated in the National Bureau of Standards Research Reactor and the neutron fluence was monitored using copper and gold foils, as well as an iron-cobalt foil.</p>			
12. KEY WORDS (Six to twelve entries; alphabetical order; capitalize only proper names; and separate key words by semicolons) fission tracks; flux monitors; glass standards; Standard Reference Material; thermal neutron irradiation; uranium.			
13. AVAILABILITY <input checked="" type="checkbox"/> Unlimited <input type="checkbox"/> For Official Distribution. Do Not Release to NTIS <input checked="" type="checkbox"/> Order From Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402. <input type="checkbox"/> Order From National Technical Information Service (NTIS), Springfield, VA. 22161		14. NO. OF PRINTED PAGES 21 15. Price	

NBS *Technical Publications*

Periodicals

Journal of Research—The Journal of Research of the National Bureau of Standards reports NBS research and development in those disciplines of the physical and engineering sciences in which the Bureau is active. These include physics, chemistry, engineering, mathematics, and computer sciences. Papers cover a broad range of subjects, with major emphasis on measurement methodology and the basic technology underlying standardization. Also included from time to time are survey articles on topics closely related to the Bureau's technical and scientific programs. As a special service to subscribers each issue contains complete citations to all recent Bureau publications in both NBS and non-NBS media. Issued six times a year.

Nonperiodicals

Monographs—Major contributions to the technical literature on various subjects related to the Bureau's scientific and technical activities.

Handbooks—Recommended codes of engineering and industrial practice (including safety codes) developed in cooperation with interested industries, professional organizations, and regulatory bodies.

Special Publications—Include proceedings of conferences sponsored by NBS, NBS annual reports, and other special publications appropriate to this grouping such as wall charts, pocket cards, and bibliographies.

Applied Mathematics Series—Mathematical tables, manuals, and studies of special interest to physicists, engineers, chemists, biologists, mathematicians, computer programmers, and others engaged in scientific and technical work.

National Standard Reference Data Series—Provides quantitative data on the physical and chemical properties of materials, compiled from the world's literature and critically evaluated. Developed under a worldwide program coordinated by NBS under the authority of the National Standard Data Act (Public Law 90-396).

NOTE: The Journal of Physical and Chemical Reference Data (JPCRD) is published quarterly for NBS by the American Chemical Society (ACS) and the American Institute of Physics (AIP). Subscriptions, reprints, and supplements are available from ACS, 1155 Sixteenth St., NW, Washington, DC 20056.

Building Science Series—Disseminates technical information developed at the Bureau on building materials, components, systems, and whole structures. The series presents research results, test methods, and performance criteria related to the structural and environmental functions and the durability and safety characteristics of building elements and systems.

Technical Notes—Studies or reports which are complete in themselves but restrictive in their treatment of a subject. Analogous to monographs but not so comprehensive in scope or definitive in treatment of the subject area. Often serve as a vehicle for final reports of work performed at NBS under the sponsorship of other government agencies.

Voluntary Product Standards—Developed under procedures published by the Department of Commerce in Part 10, Title 15, of the Code of Federal Regulations. The standards establish nationally recognized requirements for products, and provide all concerned interests with a basis for common understanding of the characteristics of the products. NBS administers this program as a supplement to the activities of the private sector standardizing organizations.

Consumer Information Series—Practical information, based on NBS research and experience, covering areas of interest to the consumer. Easily understandable language and illustrations provide useful background knowledge for shopping in today's technological marketplace.

Order the above NBS publications from: Superintendent of Documents, Government Printing Office, Washington, DC 20402.

Order the following NBS publications—FIPS and NBSIR's—from the National Technical Information Service, Springfield, VA 22161.

Federal Information Processing Standards Publications (FIPS PUB)—Publications in this series collectively constitute the Federal Information Processing Standards Register. The Register serves as the official source of information in the Federal Government regarding standards issued by NBS pursuant to the Federal Property and Administrative Services Act of 1949 as amended, Public Law 89-306 (79 Stat. 1127), and as implemented by Executive Order 11717 (38 FR 12315, dated May 11, 1973) and Part 6 of Title 15 CFR (Code of Federal Regulations).

NBS Interagency Reports (NBSIR)—A special series of interim or final reports on work performed by NBS for outside sponsors (both government and non-government). In general, initial distribution is handled by the sponsor; public distribution is by the National Technical Information Service, Springfield, VA 22161, in paper copy or microfiche form.

U.S. Department of Commerce
National Bureau of Standards
Gaithersburg, MD 20899

Official Business
Penalty for Private Use \$300