



**MANUFACTURING & QUALITY AUDIT DIVISION**  
CENTRE FOR DRUG ADMINISTRATION

Ref. No.: GUIDE-MQA-012B-004

Effective Date : 06 SEP 2004

**GUIDANCE NOTES ON  
ANALYTICAL METHOD VALIDATION :  
METHODOLOGY**

© HSA September 2004

Reproduction prohibited for commercial  
purposes.

Reproduction for internal use is authorised,  
provided the source is acknowledged.

## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

### **PREFIX**

This document is extracted from the ICH Topic Q2B, Validation of Analytical Procedures: Methodology. It is to supplement the Guidance Notes on Analytical Method Validation (GUIDE-MQA-012A).

### **INTRODUCTION**

This document is complementary to the parent guideline (ICH Harmonised Tripartite Guideline: Validation of Analytical Methods: Definitions and Terminology, ICH Topic Q2A) which presents a discussion of the characteristics that should be considered during the validation of analytical procedures. Its purpose is to provide some guidance and recommendations on how to consider the various validation characteristics for each analytical procedure. In some cases (for example, demonstration of specificity), the overall capabilities of a number of analytical procedures in combination may be investigated in order to ensure the quality of the active substance or medicinal product. In addition, the document provides an indication of the data that should be presented in an application for marketing authorisation.

All relevant data collected during validation and formulae used for calculating validation characteristics should be submitted and discussed as appropriate.

Approaches other than those set forth in this guideline may be applicable and acceptable. It is the responsibility of the applicant to choose the validation procedure and protocol most suitable for the product. However, it is important to remember that the main objective of validation of an analytical procedure is to demonstrate that the procedure is suitable for its intended purpose. Due to their complex nature, analytical procedures for biological and biotechnological products in some cases may be approached differently than in this document.

Well-characterized reference materials, with documented purity, should be used throughout the validation study. The degree of purity necessary depends on the intended use.

In accordance with the parent guideline, and for the sake of clarity, this document considers the various validation characteristics in distinct sections. The arrangement of these sections reflects the process by which an analytical procedure may be developed and evaluated.

## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

In practice, it is usually possible to design the experimental work so that the appropriate validation characteristics can be considered simultaneously to provide a sound, overall knowledge of the capabilities of the analytical procedure, for instance: specificity, linearity, range, accuracy and precision.

### **1. Specificity**

An investigation of specificity should be conducted during the validation of identification tests, the determination of impurities and the assay. The procedures used to demonstrate specificity will depend on the intended objective of the analytical procedure.

It is not always possible to demonstrate that an analytical procedure is specific for a particular analyte (complete discrimination). In this case, a combination of two or more analytical procedures is recommended to achieve the necessary level of discrimination.

#### **1.1 Identification**

Suitable identification tests should be able to discriminate between compounds of closely related structures which are likely to be present. The discrimination of a procedure may be confirmed by obtaining positive results (perhaps by comparison with a known reference material) from samples containing the analyte, coupled with negative results from samples which do not contain the analyte. In addition, the identification test may be applied to materials structurally similar to or closely related to the analyte to confirm that a positive response is not obtained. The choice of such potentially interfering materials should be based on sensible scientific judgment with a consideration of the interferences that could occur.

#### **1.2 Assay and Impurity Test(s)**

For chromatographic procedures, representative chromatograms should be used to demonstrate specificity, and individual components should be appropriately labeled. Similar considerations should be given to other separation techniques.

Critical separations in chromatography should be investigated at an appropriate level. For critical separations, specificity can be

## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

demonstrated by the resolution of the two components which elute closest to each other.

In cases where a nonspecific assay is used, other supporting analytical procedures should be used to demonstrate overall specificity. For example, where a titration is adopted to assay the drug substance for release, the combination of the assay and a suitable test for impurities can be used.

The approach is similar for both assay and impurity tests.

### **1.2.1 Discrimination of analytes where impurities are available**

For the assay, this should involve demonstration of the discrimination of the analyte in the presence of impurities and/or excipients; practically, this can be done by spiking pure substances (drug substance or drug product) with appropriate levels of impurities and/or excipients and demonstrating that the assay result is unaffected by the presence of these materials (by comparison with the assay result obtained on unspiked samples).

For the impurity test, the discrimination may be established by spiking active substance or product with appropriate levels of impurities and demonstrating the separation of these impurities individually and/or from other components in the sample matrix.

### **1.2.2 Discrimination of the analyte where impurities are not available**

If impurity or degradation product standards are unavailable, specificity may be demonstrated by comparing the test results of samples containing impurities or degradation products to a second well-characterized procedure, e.g., pharmacopoeial method or other validated analytical procedure (independent procedure). As appropriate, this should include samples stored under relevant stress conditions: light, heat, humidity, acid/base hydrolysis and oxidation.

- for the assay, the two results should be compared.
- for the impurity tests, the impurity profiles should be compared.

## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

Peak purity tests may be useful to show that the analyte chromatographic peak is not attributable to more than one component (e.g., diode array, mass spectrometry).

### **2. Linearity**

A linear relationship should be evaluated across the range (see section 3) of the analytical procedure. It may be demonstrated directly on the active substance (by dilution of a standard stock solution) and/or separate weighings of synthetic mixtures of the product components, using the proposed procedure. The latter aspect can be studied during investigation of the range.

Linearity should be evaluated by visual inspection of a plot of signals as a function of analyte concentration or content. If there is a linear relationship, test results should be evaluated by appropriate statistical methods, for example, by calculation of a regression line by the method of least squares. In some cases, to obtain linearity between assays and sample concentrations, the test data may need to be subjected to a mathematical transformation prior to the regression analysis. Data from the regression line itself may be helpful to provide mathematical estimates of the degree of linearity.

The correlation coefficient, y-intercept, slope of the regression line, and residual sum of squares should be submitted. A plot of the data should be included. In addition, an analysis of the deviation of the actual data points from the regression line may also be helpful for evaluating linearity.

Some analytical procedures, such as immunoassays, do not demonstrate linearity after any transformation. In this case, the analytical response should be described by an appropriate function of the concentration (amount) of an analyte in a sample.

For the establishment of linearity, a minimum of five concentrations is recommended. Other approaches should be justified.

### **3. Range**

The specified range is normally derived from linearity studies and depends on the intended application of the procedure. It is established by confirming that the analytical procedure provides an acceptable degree of linearity, accuracy,

## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

and precision when applied to samples containing amounts of analyte within or at the extremes of the specified range of the analytical procedure.

The following minimum specified ranges should be considered.

- for the assay of an active substance or a finished product: normally from 80 to 120 percent of the test concentration;
- for content uniformity, covering a minimum of 70 to 130 percent of the test concentration, unless a wider, more appropriate range, based on the nature of the dosage form (e.g., metered dose inhalers), is justified;
- For dissolution testing: +/-20 percent over the specified range; e.g., if the specifications for a controlled released product cover a region from 20%, after 1 hour, up to 90%, after 24 hours, the validated range would be 0-110% of the label claim.
- For the determination of an impurity: from the reporting level of an impurity to 120% of the specification; for impurities known to be unusually potent or to produce toxic or unexpected pharmacological effects, the detection/quantitation limit should be commensurate with the level at which the impurities must be controlled.  
Note: For validation of impurity test procedures carried out during development, it may be necessary to consider the range around a suggested (probable) limit;
- if assay and purity are performed together as one test and only a 100% standard is used, linearity should cover the range from the reporting level of the impurities to 120 percent of the assay specification.

#### **4. Accuracy**

Accuracy should be established across the specified range of the analytical procedure.

##### **4.1 Assay**

###### **4.1.1 Active substance**

Several methods of determining accuracy are available.

- (a) application of an analytical procedure to an analyte of known purity (e.g., reference material);
- (b) comparison of the results of the proposed analytical procedure with those of a second well-characterized

## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

procedure, the accuracy of which is stated and/or defined (independent procedure, see 1.2.2);

- (c) accuracy may be inferred once precision, linearity, and specificity have been established.

### **4.1.2 Medicinal Product**

Several methods for determining accuracy are available:

- (a) application of the analytical procedure to synthetic mixtures of the product components to which known quantities of the substance to be analysed have been added;
- (b) in cases where it is impossible to obtain samples of all drug product components, it may be acceptable either to add known quantities of the analyte to the product or to compare the results obtained from a second, well-characterised procedure, the accuracy of which is stated and/or defined (independent procedure, see 1.2.2).
- (c) accuracy may be inferred once precision, linearity, and specificity have been established.

### **4.2 Impurities (Quantitation)**

Accuracy should be assessed on samples (substance/product) spiked with known amounts of impurities.

In cases where it is impossible to obtain samples of certain impurities and/or degradation products, it is considered acceptable to compare results obtained by an independent procedure (see 1.2.2). The response factor of the drug substance can be used.

It should be clear how the individual or total impurities are to be determined, e.g., weight/weight or area percent, in all cases with respect to the major analyte.

### **4.3 Recommended Data**

Accuracy should be assessed using a minimum of 9 determinations over a minimum of 3 concentration levels covering the specified range (e.g., 3 concentrations/3 replicates each of the total analytical procedure).

## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

Accuracy should be reported as percent recovery by the assay of known added amount of analyte in the sample or as the difference between the mean and the accepted true value together with the confidence intervals.

### **5. Precision**

Validation of tests for assay and for quantitative determination of impurities includes an investigation of precision.

#### **5.1 Repeatability**

Repeatability should be assessed using:

- (a) a minimum of 9 determinations covering the specified range for the procedure (e.g., 3 concentrations/3 replicates each); or
- (b) a minimum of 6 determinations at 100% of the test concentration.

#### **5.2 Intermediate Precision**

The extent to which intermediate precision should be established depends on the circumstances under which the procedure is intended to be used. The applicant should establish the effects of random events on the precision of the analytical procedure. Typical variations to be studied include days, analysts, equipment, etc. It is not considered necessary to study these effects individually. The use of an experimental design (matrix) is encouraged.

#### **5.3 Reproducibility**

Reproducibility is assessed by means of an inter-laboratory trial. Reproducibility should be considered in case of the standardization of an analytical procedure, for instance, for inclusion of procedures in pharmacopoeias. These data are not part of the marketing authorization dossier.

#### **5.4 Recommended Data**

The standard deviation, relative standard deviation (coefficient of variation) and confidence interval should be reported for each type of precision investigated.



## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

### **6. Detection Limit**

Several approaches for determining the detection limit are possible, depending on whether the procedure is non-instrumental or instrumental. Approaches other than those listed below may be acceptable.

#### **6.1 Based on Visual Evaluation**

Visual evaluation may be used for non-instrumental methods but may also be used with instrumental methods.

The detection limit is determined by the analysis of samples with known concentrations of analyte and by establishing the minimum level at which the analyte can be reliably detected.

#### **6.2 Based on Signal-to-Noise**

This approach can only be applied to analytical procedures which exhibit baseline noise.

Determination of the signal-to-noise ratio is performed by comparing measured signals from samples with known low concentrations of analyte with those of blank samples and establishing the minimum concentration at which the analyte can be reliably detected. A signal-to-noise ratio between 3 or 2:1 is generally considered acceptable for estimating the detection limit.

#### **6.3 Based on the Standard Deviation of the Response and the Slope**

The detection limit (DL) may be expressed as:

$$DL = \frac{3.3 \sigma}{S}$$

where  $\sigma$  = the standard deviation of the response  
S = the slope of the calibration curve

The slope S may be estimated from the calibration curve of the analyte. The estimate of  $\sigma$  may be carried out in a variety of ways, for example:

## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

### **6.3.1 Based on the standard deviation of the blank**

Measurement of the magnitude of analytical background response is performed by analysing an appropriate number of blank samples and calculating the standard deviation of these responses.

### **6.3.2 Based on the calibration curve**

A specific calibration curve should be studied using samples containing an analyte in the range of DL. The residual standard deviation of a regression line or the standard deviation of y-intercepts of regression lines may be used as the standard deviation.

### **6.4 Recommended Data**

The detection limit and the method used for determining the detection limit should be presented. If DL is determined based on visual evaluation or based on signal to noise ratio, the presentation of the relevant chromatograms is considered acceptable for justification.

In cases where an estimated value for the detection limit is obtained by calculation or extrapolation, this estimate may subsequently be validated by the independent analysis of a suitable number of samples known to be near or prepared at the detection limit.

## **7. Quantitation Limit**

Several approaches for determining the quantitation limit are possible, depending on whether the procedure is non-instrumental or instrumental. Approaches other than those listed below may be acceptable.

### **7.1 Based on Visual Evaluation**

Visual evaluation may be used for non-instrumental methods, but may also be used with instrumental methods.

The quantitation limit is generally determined by the analysis of samples with known concentrations of analyte and by establishing the

## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

minimum level at which the analyte can be quantified with acceptable accuracy and precision.

### **7.2 Based on Signal-to-Noise Approach**

This approach can only be applied to analytical procedures that exhibit baseline noise.

Determination of the signal-to-noise ratio is performed by comparing measured signals from samples with known low concentrations of analyte with those of blank samples and by establishing the minimum concentration at which the analyte can be reliably quantified.

A typical signal-to-noise ratio is 10:1.

### **7.3 Based on the Standard Deviation of the Response and the Slope**

The quantitation limit (QL) may be expressed as:

$$QL = \frac{10 \sigma}{S}$$

where  $\sigma$  = the standard deviation of the responses  
S = the slope of the calibration curve

The slope S may be estimated from the calibration curve of the analyte. The estimate of  $\sigma$  may be carried out in a variety of ways including:

#### **7.3.1 Based on standard deviation of the blank**

Measurement of the magnitude of analytical background response is performed by analyzing an appropriate number of blank samples and calculating the standard deviation of these responses.

#### **7.3.2 Based on the calibration curve**

A specific calibration curve should be studied using samples containing an analyte in the range of QL. The residual standard deviation of a regression line or the standard deviation of y-

## **GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

intercepts of regression lines may be used as the standard deviation.

### **7.4 Recommended Data**

The quantitation limit and the method used for determining the quantitation limit should be presented.

The limit should be subsequently validated by the analysis of a suitable number of samples known to be near or prepared at the quantitation limit.

## **8. Robustness**

The evaluation of robustness should be considered during the development phase and depends on the type of procedure under study. It should show the reliability of an analysis with respect to deliberate variations in method parameters.

If measurements are susceptible to variations in analytical conditions, the analytical conditions should be suitably controlled or a precautionary statement should be included in the procedure. One consequence of the evaluation of robustness should be that a series of system suitability parameters (e.g., resolution test) is established to ensure that the validity of the analytical procedure is maintained whenever used.

Examples of typical variations are:

- stability of analytical solutions
- extraction time

In the case of liquid chromatography, examples of typical variations are:

- influence of variations of pH in a mobile phase,
- influence of variations in mobile phase composition,
- different columns (different lots and/or suppliers),
- temperature,
- flow rate.

In the case of gas-chromatography, examples of typical variations are:

- different columns (different lots and/or suppliers)

**GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

- temperature,
- flow rate.

**9. System Suitability Testing**

System suitability testing is an integral part of many analytical procedures. The tests are based on the concept that the equipment, electronics, analytical operations, and samples to be analyzed constitute an integral system that can be evaluated as such. System suitability test parameters to be established for a particular procedure depend on the type of procedure being validated. See pharmacopoeias for additional information.

**END OF DOCUMENT**



**MANUFACTURING & QUALITY AUDIT DIVISION**  
CENTRE FOR DRUG ADMINISTRATION

Ref. No.: GUIDE-MQA-012B-004

Effective Date : 06 SEP 2004

**GUIDANCE NOTES**

Subject: ANALYTICAL METHOD VALIDATION : METHODOLOGY

**REVISION HISTORY**

Rev No.	DCC Ref No.	Author	Date of Revision	Remarks of Revision
001	-	-	01 Nov 2001	New document.
002	-	-	09 Dec 2002	Update document format.
003	217	Terrance Ong	01 Mar 2004	1) Change 3D HSA logo to 2D HSA logo. 2) Change Centre name to Centre for Drug Administration 3) Detached Revision History from Guidance Notes.
004	265	Terrance Ong	06 Sep 2004	1) Re-located Revision History to the main document.