All data taken at Pacific Northwest National Laboratory (PNNL)

<u>Operators:</u> Jerome C. Birnbaum, Tyler O. Danby, Timothy J. Johnson, Molly Rose K. Kelly-Gorham, Rodica Lindenmaier, Tanya L. Myers

SAMPLE CONDITIONS & PHYSICAL PROPERTIES

Chemical name Sodium succinate dibasic hexahydrate

Chemical formula $C_4H_4Na_2O_4$ •6 H_2O

Synonyms Butanedioic acid disodium salt; Succinic acid disodium salt; Disodium

succinate hexahydrate

CAS number 6106-21-4

Location of field sample n/a History of sample n/a

 $\begin{array}{ll} \mbox{Molecular Weight} & 270.14 \ \mbox{g/mol} \\ \mbox{Melting Point} & > 300 \ \mbox{°C} \end{array}$

Boiling Point n/a
Density (25 °C) n/a
Hardness, Mohs scale n/a

Crystallography:

Cell dimension $a = \mathring{A} b = \mathring{A} c = \mathring{A}$

Crystal system

H-M symbol (point gr)

Space group

H-M symbol (space gr)

Crystal habit

Stated purity $\geq 99\%$

Date packed 27 April 2016 Weight: 1.912 grams

Synthesis method n/a
Synthesis reference n/a

Texture Crystalline powder

 $\begin{array}{lll} \mbox{Physical state} & \mbox{Solid} \\ \mbox{Surface roughness} & \mbox{n/a} \\ \mbox{Elemental composition} & \mbox{n/a} \\ \mbox{Isotopic composition} & \mbox{n/a} \\ \mbox{Moisture content} & \mbox{n/a} \\ \mbox{Temperature of sample} & 25 \pm 2 \ ^{\circ} \mbox{C} \\ \mbox{Substrate} & \mbox{n/a} \\ \end{array}$

INSTRUMENT PARAMETERS

Tensor 37 FT-IR manufactured by Bruker Optics

External diffuse reflectance accessory A 562-G integrating sphere

Sphere diameter 75 mm Angle to normal incidence 14.8°

Sphere opening diameter 19 mm (entrance port)

Spectral range 7,500 to 600 cm⁻¹ saved; 7500 to 600 cm⁻¹ reported

Beamsplitter Ge on KBr

Detector (dia. Det. Port in sphere) 2×2 mm, 60° field of view MCT (550; 0.9); 1 cm

Apodization function Blackman-Harris 3-term

Aperture 6 mm

Coadded scans 2048

Scanner speed 40 kHz

Switch gain on 512 points

Low pass filter Open

Scan technique double-sided, forward-backward

Non-linear correction On


High and low folding limit 15800.54-0.00 cm⁻¹

Phase resolution 32.00Phase correction mode MertzZerofilling $4 \times$

Wavenumber accuracy $\pm 0.4 \text{ cm}^{-1}$ Spectral resolution 4 cm^{-1}

Accuracy verification 10/28/2015

Wavelength vetted on: ICL polystyrene standard #0009-7394-0025A, thin film Reflectance: ±2% using SRS reflectance standards 50-010-DH27B-4878






Figure 1: The Bruker 562-G integrating sphere (a) and Tensor 37 (b)

Photographs of sample Sodium Succinate Dibasic Hexahydrate



Figure 2: Sodium Succinate Dibasic Hexahydrate in Sigma-Aldrich container.



Figure 3: Sodium Succinate Dibasic Hexahydrate loaded in IR sample cup.

PARTICLE SIZE PREPARATION AND CHARACTERIZATION

Optical microscopy —

A Keyence VHX-1000 digital microscope with 16-bit resolution is used to provide photomicrographs of the various samples and particle sizes. Software included with the microscope differentiates the brightness and colors in the image and extracts the bright objects to produce a binary image. The software assumes all adjacent bright points are part of the same object then calculates the area for each of these objects. The area (A) is used to calculate the mean particle diameter (d) by assuming the particles are spherical and using the relationship $d=(4*A/\pi)1/2$. Although the assumption of spherical particles is clearly not always valid, this procedure provides a reasonable estimate of the mean particle size.




Figure 4: Photomicrograph of Sodium Succinate Dibasic Hexahydrate.




Figure 5: Particle size distribution of Sodium Succinate Dibasic Hexahydrate.